

Certificate of Analysis

Company: Theory Wellness of VT
 768 Putney Rd
 Brattleboro, VT 05301

Sample ID: Tincture - CBD - Liquid Bulk

Lot: 0054-TIN600C1

Report Date: 1/15/2024

Date Analyzed: 1/12/2024

Matrix: Tincture

Analyst: 011

Customer ID: 230609-0

Date Sampled: N/A

Report ID: C240104BK

Grower License #: MANU0054

Date Received: 1/4/2024

Cannabinoid Summary

Cannabinoid Profile	LOQ (mg/g)	Concentration (mg/g)	Weight (%)
CBDVA	0.0005	<LOQ	<LOQ
CBDV	0.0012	<LOQ	<LOQ
CBDA	0.0008	<LOQ	<LOQ
CBGA	0.0008	<LOQ	<LOQ
CBG	0.0019	<LOQ	<LOQ
CBD	0.0019	16.18	1.62
THCV	0.0021	<LOQ	<LOQ
CBN	0.0013	<LOQ	<LOQ
Δ9-THC	0.0020	1.12	0.11
Δ8-THC	0.0019	<LOQ	<LOQ
THC-A	0.0034	<LOQ	<LOQ
CBC	0.0024	<LOQ	<LOQ
Total THC		1.12	0.11
Total CBD		16.18	1.62
Total Cannabinoids		17.30	1.73

0.11%

Total THC

1.62%

Total CBD

1.73%

**Total
Cannabinoids**

0.11%

Δ9-THC

N/A

**Percent
Moisture**

1 : 14.4

**THC : CBD
Ratio**

Cannabinoids Methodology: High Performance Liquid Chromatography (HPLC) using PerkinElmer FLEXAR™ with Photo Diode Array Detector (PDA)

Total CBD and total THC are calculated values, to account for assumed decarboxylation from the acid form (THCA or CBDA) to the neutral form, causing weight loss of the acid group. These values are calculated as follows:

Total THC = (THCA x 0.877) + Δ9-THC

Total CBD = (CBDA x 0.877) + CBD

Ratio of Total CBD: Total THC

Reagent Blanks: < LOQs for all analytes

LOQ = The lowest quantity that this method can reliably detect. Any cannabinoid that was not detected is assumed to be less than the stated LOQ (<LOQ).

All results reflect dry weight of material, based on % moisture of the sample.

Measurement of Uncertainty (MU): the parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the particular quantity subject to measurement.

Δ9-THC MU = ±0.005%

Total THC MU = ±0.007%

All other cannabinoid MU values are available upon request.

All moisture analysis is determined by loss-on-drying measurement using OHAUS Model MB90 Moisture Content Readers.

This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context. Results apply to the samples as received.

Certified by:

Luke E. M.

Luke Emerson Mason (Laboratory Director, Bia Diagnostics)

